CERTIFICATE OF ANALYSIS

Complement Technology, Inc.
4801 Troup Hwy, Suite 701
Tyler, Texas 75703, USA

Product: **VBS^{o}**
Catalog # **B114**
Prep. Date: _____

Specifications	**Limits**	**Results**
FILL VOLUME	250 – 275 mL	260 mL
BUFFER	Veronal Buffered saline without calcium and magnesium, pH 7.3 ± 0.1	Veronal Buffered saline without calcium and magnesium, pH 7.3 ± 0.1
PRESERVATIVE	Sodium Azide, 0.025%	Sodium Azide, 0.025%
FILTRATION	0.45 μm filter	0.45 μm filter
IONIC STRENGTH	14 mmho ± 1 @ 22 ºC	14 mmho ± 1 @ 22 ºC

STORAGE CONDITIONS

Buffer may be stored for 90 days at 4 ºC.
Buffer may be stored for 4 years at -20 ºC. If frozen, buffer should be warmed and stirred prior to use. Buffer should then be chilled to 4 ºC before use.

FOR RESEARCH USE ONLY
NOT FOR HUMAN OR DRUG USE

_____ _____
Signature of Analyst Date of Analysis

Phone: 1-903-581-8284 FAX: 1-903-581-0491 Email: contactCTI@aol.com
Web Site: www.ComplementTech.com